

# EC Declaration of Conformity

Manufacturer: Wenzhou Zuohan Clothing Co., Ltd.  
Address: No.696 Binhai Fifth Road, Wenzhou Economic And  
Technological Development Zone, Zhejiang, China

European Representative: BTOP GLOBAL GMBH  
DUISBURGER STR.3A 41460 NEUSS GERMANY

Product Name: KN95 Disposable Mask  
Model No: HY-95  
Classifications: FFP2  
Test Standard: EN 149:2001+A1:2009 Respiratory protective devices – Filtering half masks  
to protect against particles – Requirements, testing, marking

We herewith declare that the above mentioned products meet the requirements of the Regulation (EU) 2016/425 Personal Protective Equipment (PPE). The submitted products have been tested and found in compliance with the following European Standards: EN149:2001+A1:2009. All supporting documentations are retained under the premises of the manufacturer.

The manufacturer is exclusively responsible for the declaration of conformity.

Place of issue: No.696 Binhai Fifth Road, Wenzhou Economic And  
Technological Development Zone, Zhejiang, China

Signature: zhou zuo

Date: April 27, 2020







WORK SHOP

# KN95 Four-layer mask -structure

## Multilayer protection



Non-woven  
50g

Hot air  
cotton50g

Meltblown cloth  
25g

Non-woven 25g

## KN95Mask-Four layer

**Frist:**

50g Non-woven fabric

Second:

50g hot air cotton

**Third:**

25g meltblown cloth

Fourth:

25g non-woven fabric

**Package:**

**20PCS/Box**

Standard:


EN 149:2001+A1:2009

Filtration rate:

**PFE ≥ 95%**



# KN95 Five-layer mask- structure



## KN95 Mask-Five layer

<b>First:</b>	50g Non-woven fabric
Second:	50g Hot air cotton
<b>Third:</b>	25g Meltblown cloth
Fourth:	25g Meltblown cloth
<b>Fifth:</b>	25g Non-woven fabric
Package:	20 pcs/box
<b>Standard:</b>	EN 149:2001+A1:2009
Filtration rate:	BFF ≥ 95%



180011112242



(2018) 国认监认字 (244) 号



中国认可  
国际互认  
检测  
TESTING  
CNAS L6780

# 检 验 报 告

## TEST REPORT



Scan is following us

Scan checks for authenticity



\* W 2 0 2 0 0 9 8 8 0 E \*

浙江省轻工业产品质量检验研究院

Zhejiang Light Industrial Products Inspection and Research Institute

国家纺织服装产品质量监督检验中心（浙江）

National Textiles and Garment Quality Supervision Inspection Center(Zhejiang)

# Test report

**Number:**W202009880E

page1/2

Name of Customer	Wenzhou Zuohan Clothing Co., Ltd	Address	NO. 696 Binhai Fifth Road, Wenzhou Economic And Technological Development Zone, Zhejiang, China
Manufacturer	---	Address	---
Sample information	Name of sample: KN95 mask Characteristics of sample: white Trademark of sample: --- Specification/model: --- Level: KN95 Category of safety specification: --- Art. No.: --- -----		
Above-mentioned information by Customer-supplied			
The sent way of sample	Courier	Sample quantity	20 pieces
Receiving Date of Sample	2020/04/24	Test Category	Entrusted inspection
Date of Testing	2020-04-24- 2020-04-29		
RatingRequirements	GB 2626-2006		
Test Summary:  See the attached page for the results.   <div style="text-align: right;">                     Test Seal                       Date of Approval: 2020-05-26                 </div>			
Remarks			

Approved by:





# Test report

Number:W202009880E

page2/2

ITEM	STANDARD (KN95)	RESULT	RATING
1. FILTER EFFICIENCY(SALT MEDIUM) (GB 2626-2006 6.3)(%)			
-	$\geq 95.0$	98.2	PASS
2. RESPIRATORY RESISTANCE (GB 2626-2006 6.5 & 6.6)(Pa)			
-INSPIRATORY RESISTANCE	$\leq 350$	106	PASS
-EXPIRATORY RESISTANCE	$\leq 250$	91	PASS
3. APPEARANCE (GB 2626-2006 6.1 & 6.2)			
-	Meet the requirements of GB 2626-2006 standard	The appearance quality of 2 masks conforms to the standard	PASS

Picture(s) of sample



—End of report—

# DECLARATION

1. Our organization guarantees impartiality, independence and honesty of inspection, and is responsible for the results of inspection, keeping the samples supplied by the entrusting party confidential and at the same time protecting the ownership of the samples supplied.
2. The test report will be deemed invalid without signatures of the inspector/reviewer and authorized personnel, and the red special inspection stamp of our organization.
3. The test report will be invalid if it is altered. Copies of the report are invalid without the red special inspection stamp of our organization.
4. The test results shown in this report are applicable only to the samples provided by customers.
5. All the pages of the report are integral parts of the report. Our organization will not be responsible for any misunderstanding or other results caused by using separate page(s) of the report.
6. If there is any dissent of the report, the entrusting party shall notify our organization timely. For the mandatory inspection given by governmental administration departments, and dissent about the sample being tested or test results on the report should be dealt with in accordance with national regulations.

## 浙江省轻工业品质量检验研究院及附设的检验中心 The Affiliated Inspection Centers

浙江省轻工业品质量检验研究院  
Zhejiang Light Industrial Products Inspection and Research Institute  
国家纺织服装产品质量监督检验中心（浙江）  
National Textiles and Garment Quality Supervision Inspection Center (Zhejiang)  
地址：浙江省杭州市江干区下沙路 300 号 6 号楼  
Address: Building No.6, 300 XiaSha Road, Hangzhou, Zhejiang

联系电话：0571-85122669  
Telephone: 0571-85122669

E-Mail: 1137907994@qq.com

国家家具产品质量监督检验中心（浙江）  
National Center for Quality Supervision Inspection of Furniture (Zhejiang)  
浙江省室内安全及家具产品质量检验中心  
Zhejiang Center of Quality Test for Indoor Safety and Furniture Products  
地址：浙江省杭州市余杭区良渚街道经一路 1 号良渚大学科技园 4 号楼  
Address: Building 4 LiangZhu University Science and Technology Park, No.1,  
JingYi Rd, Yuhang District, Hangzhou, Zhejiang

联系电话：0571-89009556  
Telephone: 0571-89009556

E-Mail: 2047699564@qq.com

浙江省轻工及五金产品质量检验中心  
Zhejiang Center of Quality Test for Light Industry and Hardware Products  
浙江省体育用品质量检验中心  
Zhejiang Center of Quality Test for Sports Products  
地址：浙江省杭州市西湖区天目山路 222 号 3 号楼  
浙江省杭州市余杭区良渚街道经一路 1 号良渚大学科技园 3 号楼  
Address: No. 222 Tianmushan Rd., Hangzhou, Zhejiang  
Building 3 LiangZhu University Science and Technology Park, No.1,  
JingYi Rd., Hangzhou, Zhejiang

联系电话：0571-89001107  
Telephone: 0571-89001107

E-Mail: wujtest@126.com

国家锁具产品质量监督检验中心（浙江）  
National Center for Quality Supervision Inspection of Lock (Zhejiang)  
浙江省锁具产品质量检验中心  
Zhejiang Center of Quality Test for Lock Products  
地址：浙江省杭州市塘苗路 24 号  
Address: No. 24 Tangmiao Rd., Hangzhou, Zhejiang

联系电话：0571-85027738  
Telephone: 0571-85027738

E-Mail: Locktest@126.com

网上业务受理/报告查询: <http://www.zjttj.cn>  
Online Business-Reception/Report Inquires: <http://www.zjttj.cn>

投诉电话：0571-85023552  
Complaint Tel: 0571-85023552





KN95 Mask




KN95 Mask





KN95 Mask



KN95 Mask



Filtration efficiency  
up to >95%



Respiratory resistance

<80pa

Filtration

95%

Formaldehyde

0%

KN95

Plastic nose bridge fits the face better  
After wearing the mask .press the plastic nose  
bridge. Make it fit the face .Prevent pollutants

0  
Glue

0  
Formaldehyde

360°  
Facial fit  
5  
Protection



soft cotton .skin-friendly

NO ALLERGIES, NO IRRITATION

Not irritating

